



**Corning HTL SA**  
4 Daniszewska Str., 03-230 Warsaw, Poland  
phone: +48 22 492 19 00  
htl.info@corning.com  
www.htl.pl



AP 169



Calibration laboratory accredited by  
Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA  
that include recognition of calibration certificates.  
Accreditation No AP 169

## CALIBRATION CERTIFICATE

Date of issue: January 23, 2024

Certificate No: SW/0104/2024

Page: 1 / 2

<b>OBJECT OF CALIBRATION</b>	Single-channel piston pipette Manufacturer: Corning HTL SA Serial number: 04F15397 Volume: ( 20 - 200 ) $\mu$ l Pipette tips: AXYGEN
<b>APPLICANT</b>	Corning HTL SA 4 Daniszewska Str., 03-230 Warsaw, Poland
<b>PLACE OF CALIBRATION</b>	Corning HTL SA 4 Daniszewska Str., 03-230 Warsaw, Poland
<b>CALIBRATION METHOD</b>	PN-EN ISO 8655-6:2022
<b>ENVIRONMENTAL CONDITIONS</b>	Air temperature: ( 22.4 $\div$ 22.9 ) $^{\circ}$ C Relative humidity: ( 50 $\div$ 56.5 ) % Atmospheric pressure: ( 1009.7 $\div$ 1010.3 ) hPa Water temperature: ( 22.4 $\div$ 22.6 ) $^{\circ}$ C
<b>DATE OF CALIBRATION</b>	January 20, 2024
<b>TRACEABILITY</b>	This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).
<b>CALIBRATION RESULTS</b>	The results have been presented on page 2 of this certificate including uncertainty of measurement. The measurement results only apply to the calibrated instrument.
<b>UNCERTAINTY OF MEASUREMENT</b>	Uncertainty of measurement has been evaluated in compliance with EA-4/02. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor $k = 2$ .



Quality Control Supervisor  
  
Krzysztof Kostro-Olechowski

This certificate may be presented or copied as a whole document only.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 169

Date of issue: January 23, 2024

Certificate No: SW/0104/2024

Page: 2 / 2

**CALIBRATION  
RESULTS**

Calibration results are the following:

The value of reference volume $V_0$ $\mu\text{l}$	Measured volume $V$ $\mu\text{l}$	Measurement error $\Delta V$ $\mu\text{l}$	Uncertainty of measurement $U$ $\mu\text{l}$
20	20.40	0.40	0.05
100	100.22	0.22	0.33
200	200.31	0.31	0.33

**ADDITIONAL  
INFORMATION**

The measured volume value is based on the reference temperature = 20 °C.

Authorized by: Renata Frączek

